



Product Service

# CERTIFICATE

No. B 11 04 47365 022

**Holder of Certificate:** PG Drives Technology Ltd.10 Airspeed Road, Christchurch  
Dorset BH23 4HD  
UNITED KINGDOM**Production Facility(ies):**

76999, 47365

**Certification Mark:****Product:**

Controller

**Model(s):**

Drive control systems for electrically driven wheelchairs, type newVSI

**Parameters:**

Nominal Battery Voltage:	24V
Operating Voltage Range:	16-35Vdc
Nominal Drive Current Limit:	Up to 50A per channel
Rated brake current:	1 A max.
Degree of protection:	IPX4

Remark:

See Test Report for approved module types.

**Tested according to:** ISO 7176-14:2008

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 71385677**Date,** 2011-05-06

Page 1 of 1



(Torsten Zimmer)

